A-GAS: a Probabilistic Approach for Generating Automated Gait Assessment Score for Cerebral Palsy Children

Rishabh Bajpai, and Deepak Joshi

Abstract—Gait disorders in children with cerebral palsy (CP) affect their mental, physical, economic, and social lives. Gait assessment is one of the essential steps of gait management. It has been widely used for clinical decision making and evaluation of different treatment outcomes. However, most of the present methods of gait assessment are subjective, less sensitive to small pathological changes, time-taking and need a great effort of an expert. This study proposes an automated, comprehensive gait assessment score (A-GAS) for gait disorders in CP. Kinematic data of 356 CP and 41 typically developing subjects is used to validate the performance of A-GAS. For the computation of A-GAS, instance abnormality index (AII) and abnormality index (AI) are computed. AII quantifies gait abnormality of a gait cycle instance, while AI quantifies gait abnormality of a joint angle profile. AII is calculated for all gait cycle instances by performing probabilistically and statistical tests. Abnormality index (AI) is a weighted sum of AII, computed for each joint angle profile. A-GAS is a weighted sum of AI, calculated for a lower limb. Moreover, a graphical representation of the gait assessment report, including AII, AI, and A-GAS is generated to understand the results better. Furthermore, the study compares A-GAS with a present rating-based gait assessment scores to understand fundamental differences between them. Finally, A-GAS's performance is verified for a high-cost multicamera set-up using nine joint angle profiles and a low-cost single camera set-up using three joint angle profiles. Results show no significant differences in performance of A-GAS for both the set-ups. Therefore, A-GAS for both the set-ups can be used interchangeably.

Index Terms—Human locomotion, gait assessment, cerebral palsy, automation

I. INTRODUCTION

ABNORMAL bone growth, bone deformation, loss or reduction of neural activation in muscles, loss of proprioception and muscle abnormalities may lead to abnormal walking patterns [1]. Cerebral Palsy (CP) is the most common pediatric neurodevelopmental disorder that affects motor skills, muscle tones, balance and movements [2]. Since CP is a non-progressive disorder, accurate and timely gait management may significantly recover the movements. Assessment of abnormal walking patterns is considered one of the most crucial steps of gait management in CP. Accurate assessment of gait helps the doctor in selecting appropriate interventions [3]. For example, in case of minor gait deviations with spasticity,

botulinum toxin may be provided for relaxing muscles and avoiding muscle contracture [4]. Then, a series of physical therapy sessions are conducted to strengthen the muscles. In between the sessions, a gait assessment is performed to track the changes in gait and formulate new management strategies. However, most of the present gait assessment tools are not very sensitive to record small pathological changes in the gait [5]. The following subsection overviews present gait assessment tools by explaining their assessment methodologies, clinical relevance, limitations and the advantage over other methods. Broadly, gait assessment tools can be classified into three categories: qualitative methods, quantitative methods, and clinical gait analysis.

A. Literature Review

Qualitative methods include both observation-based methods and questionnaire-based methods. The fundamental principle of observation-based methods is to detect and understand the abnormality in the gait, mostly by visualizing the video of gait. Therefore, these methods demand a great understanding of the biomechanics of walking and experience in gait assessment. The observer generates a report, and calculates a gait assessment score based on the observations of gait's basic parameters such as stride, joint angles, posture and movements. Most of the popular observation-based classifications systems are based on work of Sutherland et al. [6], Winter et al. [7] and Rodda et al. [8]. On the other hand, questionnaire-based methods need less understanding of walking and no experience in gait assessment. These methods consist of questions related to walking and other movements of the lower limbs. Gillette Functional Assessment Walking Scale is a popular 10-point scale parent-report questionnaire that can be used with no or little knowledge of gait [9]. Functional Mobility Scale [10], Gross Motor Function Classification System (GMFCS) [11], Pediatric Quality of Life inventory [12], Pediatric Evaluation of Disability Inventory [13], Gait Outcomes Assessment List [14] and Activities Scale for Kids [15] are some other widely used questionnaires-based assessments. No doubt, these methods are beneficial for preliminary gait assessment in the absence of an expert. However, they are less accurate, less comprehensive, and biased towards some common gait pathologies.

Most of the shortcomings of qualitative methods are addressed by quantitative methods. In these methods, gait data, mostly kinematics of joints and segments of a patient's lower limb, are

R. Bajpai and D. Joshi are with Indian Institute of Technology Delhi, 110016, India, and All India Institute of Medical Science Delhi, New Delhi, 110029, India. E-mail: {bmz208129@iitd.ac.in, rishabh@iitd.ac.in, joshid@iitd.ac.in}.

Manuscript received ;

compared with normal kinematics data for understanding areas of abnormality in the gait. The two most common quantitative approaches involve manual assessment score calculation and automatic computation. In manual calculation-based methods, an expert (also known as rater) rates gait-related parameters by comparing them with their corresponding normal parameters values. For example, Edinburgh Visual Gait Score [16] was developed for assessing the gait of CP children in all three planes. It uses 17 gait parameters of ankle, knee, hip, pelvis and trunk covering all three planes. It uses three-point scale for each parameter, where '0' denotes normal, '1' denotes moderate deviation and '2' denotes marked deviation. After calculating ratings for all parameters, a gross rating is calculated by summing the individual ratings. As a results, the range of gross rating's value is 0-34. One key limitation of this method is its subjectivity, due to which the ratings in this method may vary with raters [17]–[19]. Observational Gait Scale [20] is a manual assessment score calculation-based method developed for assessing the gait of CP children in all three planes. It is based on twenty-six parameters (features) of gait, related to ankle/foot, knee, hip, pelvis joints. Physician's Rating Scale [21] and Salford Gait Tool [22] are two other popular scores to assess ankle, knee and hip joints. Physician's Rating Scale has many variants, namely, modified PRS [21], Abbreviated PRS [23], Visual Gait Analysis Scale [24] and observational gait scale [25]. All are based on a scoring system with a maximum score ranging from 1 to 26. Salford Gait Tool [22] is a five-point scale that assesses three joints at six gait events: initial contact, end of double support, midstance, start of double support, toe-off, and mid-swing. Manual calculation-based methods are highly useful in clinical settings. However, there are some major limitations of these methods discussed by many researchers in the last few decades [5], [17]–[19], [26]. They need an experienced rater, they are highly subjective, false-negative reports are very common due to the use of limited gait-related parameters (features), time-consuming, less sensitive to small changes in the gait.

On the other hand, automatic computational methods are objective because they generally use dimension reduction of the gait data for computing gait assessment score and classifying gait into normal and abnormal gait. Recently, Lei Wang et al. [27] proposed an inertial measurement unit (IMU) based gait normalcy index. The normalcy index was calculated from six kinematic parameters and six three spatial-temporal parameters extracted from two shank-mounted IMUs [27]. IMU-based methods can accurately extract a smaller number of gait pathology relevant parameters. This limits the usability of IMU-based methods in clinical gait analysis. Similarly, Jing Gao et al. [28] proposed an IMU-based abnormal gait recognition algorithm. Their algorithm uses a long short-term memory network and convolutional neural network for detecting tiptoe, hemiplegic, and cross-threshold gait abnormal gait. Their method is cost-efficient and can be used for the recognition of a set of abnormal gaits. However, every patient has a unique gait abnormality [1], and it is not feasible to train a neural network for all possible gait disorders. Trong-Nguyen Nguyen and Jean Meunier [29] proposed an approach that estimates a human walking gait abnormality index using an

adversarial auto-encoder. They fit 3-D cloud data to cylinders and fed these cylinders to an adversarial auto-encoder to calculate some potential gait abnormality indices. In their other work [30], they tried to estimate skeleton-based gait abnormality index by sparse deep auto-encoder. Xiang Chen et al. [31] proposed a three IMUs based objective motion disability tool of CP patients. They compared gait acceleration data from three IMUs placed on the trunk and thighs of healthy and CP subjects to calculate a global score. The field of use of automatic computational methods is in its embryo stage. Present methods either provide a global abnormality index or perform classification into a small number of gait abnormalities. Most of the automatic computational methods are less intuitive, they use black-box signal processing due to which the localization of gait abnormality from computed gait score becomes impossible. Their inability to provide a comprehensive assessment report with localization of abnormal events in the gait is the primary reason for their less clinical acceptance.

Clinical gait analysis is considered the most comprehensive gait assessment procedure. It may include the processing of various gait-related signals like three-dimensional motion data (force, kinematic, kinetic of joints), electromyography (EMG), physical examination, CT scan and magnetic resonance imaging (MRI) [32]. This approach is time-consuming, needs a team of experts from various domains, subjective and detailed. It is recommended mainly in severe cases of gait abnormality [32].

B. Motivation and Contributions

As discussed above, a large literature can be found on the verification and modification of these methods. No doubt, these methods can explain gait abnormalities in most cases. However, there are four major limitations of these methods. First, most of these methods need significant efforts of an experienced evaluator. Second, they are subjective and prone to misclassification. Third, most of them use features of gait for assessment which make them are less sensitive to detect minor changes in gait and capture uncommon gait abnormalities. Fourth, they are less user-friendly as they provide either expanded raw data which difficult to interpret or provide some global indices, which makes the process of localization of abnormality more difficult. Many studies highlighted the shortcomings of the present gait assessment method [17], [33]–[35]. Several studies reported large variability of assessment score when a single subject has been analysed in a number of different laboratories [36], [37]. However, these methods are being used in clinical settings due to their popularity, intuitiveness and unavailability of a comprehensive automated gait assessment system. After understanding the merits, demerits of the present methods of gait assessment and needs of the clinicians, we concluded the following expectation from a gait assessment tool. The gait assessment should:

- be objective, comprehensive, easily implementable and user-friendly.
- use easy mathematics and non-black-box signal processing for better understanding for doctors/clinicians.

- use a continuous valued scale for better sensitivity to minor changes in gait patterns.
- hold completeness property. By considering all joints and gait cycle instances completeness of gait assessment can be assured.
- give freedom to the user to manipulate parameters used for gait assessment.
- be fast. By using present methods, each session of gait assessment and interpretation of the results by experts takes three to six hours [38].

This study tried to address above mentioned needs, and contributions of the study are as follows:

Contributions:

- 1) Automated-Gait Assessment Score (A-GAS) is the first subjective comprehensive gait assessment methodology developed for CP children (as per the authors' knowledge).
- 2) There are three components of A-GAS report corresponding to different levels of gait analysis. The first component depicts the global abnormality index of a limb, the second component depicts the abnormality index of each joint, and the third component depicts the abnormality of each gait cycle instance of each joint.
- 3) The methodology behind the algorithm of A-GAS resembles the thinking process of a clinician while comparing an abnormal gait with a normal gait. Therefore, it is easy to understand and manipulate and potentially provide better clinical acceptance.
- 4) The study tested performance of A-GAS on two joint angle profile configurations. Both the configurations can be used with the present instrumentations. The first configuration targets gait laboratories with multiple camera setup. It includes nine joint angle profiles viz., ankle sagittal, knee sagittal, hip sagittal, pelvic sagittal, hip transverse, pelvic transverse, hip coronal, pelvic coronal and foot transverse. The second configuration targets gait laboratories with Kinect, one camera set-up or goniometers. It includes three joint angle profiles viz., ankle sagittal, knee sagittal and hip sagittal.

The rest of the work is as follows: step to compute A-GAS are described in Section II; performance of the proposed assessment system with a comparison with one rater-based method are presented in Section III; the significance of the research, limitations and future plans of A-GAS are discussed in Sections IV; finally, the conclusion of the work is drawn in Section V.

II. METHODOLOGY

A. Dataset Description

This study includes rich datasets for testing the performance of A-GAS on the many possible gait abnormalities. One CP children's and two typically developing (TD) children's, publicly available datasets with a total of 397 subjects are used.

- 1) CP children dataset [39]: The dataset was recorded by the clinical motion analysis laboratory of University Hospital Pellenberg with approval of the medical ethics

committee of University Hospitals Leuven [40], [41]. It consists of gait analysis sessions recorded from 356 children (age 4 to 18 years) with unilateral or bilateral spastic CP (GMFCS level I-III). For this study, only kinematics data from 1719 trials recorded by the Vicon system (Vicon Motion Systems Ltd., Oxford, UK) at self-selected speed is used. Each trial is a gait cycle with 51 number of samples. A sample plot of distribution of knee angle's profile is shown in Fig. 2.A.

- 2) Normal dataset [42], [43]: The dataset was recorded with signed informed consents of the participants and approval of the Institutional Research Ethics Committee of Fondazione Don Carlo Gnocchi. It consists of data from 41 TD children (20 from [42], 21 from [43]) aged 6 to 17 years. This dataset contain less number of subjects when compared to CP dataset. However, it is found to be correlated with another dataset of TD with a similar age group. For this study, only kinematics data recorded by a 9-camera motion capture system (SMART system, BTS, Garbagnate Milanese, Italy) at self-selected speed is used. For further processing, each trial of kinematic data is down-sampled from 101 to 51 gait cycle instances to match with CP dataset's sampling frequency. A sample plot of distribution of knee angle's profile is shown in Fig. 2.A.

Ankle sagittal, knee sagittal, hip sagittal, pelvic sagittal, hip transverse, pelvic transverse, hip coronal, pelvic coronal and foot transverse joint angle profile is extracted from the datasets (see Fig. 1).

Fig. 1. Graphics illustrating three orthogonal planes with respect to human body and five segments/joints of the lower portion of human body. Movement of these segments/joints in the three planes during gait generate joint angle profiles.

B. Normal vs abnormal comparison

For the feasibility of using a probabilistic approach for gait assessment, the assumption that gait angle profiles of CP significantly differ from TD must hold. To verify this assumption, some statistical tests are performed on each joint profile. First, the normality of histograms for CP gait cycles and TD gait cycles at each cycle instance is verified by Shapiro-Wilk test. Then normal probability density functions (PDFs) are fitted to the data at each cycle instance with 1719 number of CP data-points and 41 number of TD data-points. A sample plot of PDFs of CP and TD angle profiles at 30th

gait cycle instance is shown in Fig. 2B. Then, the significance of difference of means of CP gait cycles and TD gait cycles at each cycle instance is evaluated by a T-test. P values obtained by T-test for knee angle profile are shown in Fig. 2D. It shows that the difference in means of CP gait cycles and TD gait cycles is significant (p value less than 0.05) at most of the gait cycle instances for knee joint angle. Similar results are obtained for other joint angle profiles. So we can conclude that the assumption that gait angle profiles of CP significantly differ from TD holds. Therefore, there is a possibility of the existence of a probabilistic solution for quantifying gait abnormality.

C. Instances weight calculation

In the previous analysis, at some gait cycle instances, the difference of means of CP gait cycles and TD gait cycles was not significant Fig. 2D. This suggests there is an absence of evidence to reject the null hypothesis statistically. In other words, we do not have the confidence to say that means of CP gait cycles and TD gait cycles at these cycle instances are different. While at most of the cycle instances, the difference was significant. This suggests, there is sufficient evidence to reject the null hypothesis statistically. In other words, we do have the confidence to say that means of CP gait cycles and TD gait cycles at these cycle instances are different. Mathematically, the confidence as mentioned above is analogous to the p-value of the T-test. The gait cycle instances with p-values are close to one, have lesser significance for distinguishing values of CP gait cycle with TD gait cycle when compared with the gait cycle instances with p-value is close to zero. As it can be observed that significance and p-values are inversely proportional. To quantify this significance, a weight instance vector is defined at all cycle instances. By using a non-linear mapping, the p-value are converted into an instance weight at each gait cycle instance. This non-linear mapping is given by the following equation 1, and its shown in Fig. 2C.

$$InstanceWeight(Pvalue) = 1 + \frac{(0.25 - 1)}{1.5^{1/Pvalue}} \quad (1)$$

Instance weight is calculated for all instances and stored in a vector 'I'. The least value of I is set to 0.5 to avoid vanishing of the value of the entity multiplied with it. The nature of the vector is chosen to be non-linear with a slower descent rate at smaller p-values values to take advantage of the physical significance of p-values.

D. Computation of A-GAS of a given sample

An abnormality score A-GAS is computed for representing the degree of abnormality in gait. Let us assume for a given sample X , angle profile of knee joint in sagittal plane at i th gait cycle instance is denoted by X_i^{kneeS} and the abnormality index of knee angle profile in sagittal plane at i th gait cycle instance is denoted by AI_i^{kneeS} . So, the expression for computing AI_i^{kneeS} is given by the following equations 2.

$$AI_i^{kneeS} = I_i^{kneeS} * L_i^{kneeS} \quad (2)$$

Where, I_i^{kneeS} is i th gait cycle instance value of instance weight vector I^{kneeS} computed for knee joint, L_i^{kneeS} is i th gait cycle instance value of modified likelihood ratio vector L^{kneeS} of knee joint and AI_i^{kneeS} is i th gait cycle instance value of instance abnormality index vector AI^{kneeS} of knee joint. For computation of L_i^{kneeS} , likelihood of X_i^{kneeS} for CP gait cycle at i th instance and TD gait cycle at i th instance are calculated from their respective fitted PDFs. To elaborate, in step 1, PDFs for CP gait cycles and TD gait cycles are computed for a gait cycle instance. A sample PDF plot is shown in Fig. 3B for gait cycle instance 30 of sample shown in Fig. 3A. In step 2, likelihood of the values of a given sample is noted. In the given example, $X_{30}^{knee} = 27$, $\mathcal{L}(\mu_{30}^{CP}, \sigma_{30}^{CP} | 27) = 0.03367$ and $\mathcal{L}(\mu_{30}^{TD}, \sigma_{30}^{TD} | 27) = 0.00830$. Here, μ_{30}^{CP} and σ_{30}^{CP} are mean and SD of CP profile value's distribution at 30th gait cycle instance, respectively. μ_{30}^{TD} and σ_{30}^{TD} are mean and SD of TD profile value's distribution at 30th gait cycle instance, respectively. $\mathcal{L}$ is the likelihood function. In step 3, likelihood ratio (LR_i^{kneeS}) of TD to CP is computed by equation 3. For the sample, likelihood ratio of TD to CP is $0.008305/0.03367 = 0.2467$.

$$LR_i^{kneeS} = \frac{\mathcal{L}(\mu_i^{TD}, \sigma_i^{TD} | X_i^{kneeS})}{\mathcal{L}(\mu_i^{CP}, \sigma_i^{CP} | X_i^{kneeS})} \quad (3)$$

In step 4, modified likelihood ratio L_i^{kneeS} is computed by the following equation 4.

$$L_i^{kneeS} = \begin{cases} 1 - LR_i^{kneeS} & \text{if } LR_i^{kneeS} \leq 1 \\ 0 & \text{if } LR_i^{kneeS} > 1 \end{cases} \quad \text{if } \mathcal{L}(\mu_i^{CP}, \sigma_i^{CP} | X_i^{kneeS}) = 0 \quad (4)$$

According to the equation 4, modified likelihood ratio of the given sample is $1 - 0.2467 = 0.7533$. Modified likelihood ratio is inspired from Bayes' discrimination rule which states that a sample x is assigned to a group that maximizes $\pi_i \mathcal{L}_i(x)$ is higher, where π_i is prior probability of group i . However, it is not possible to accurately estimate π for CP and TD group at each gait cycle instance. Therefore, it is assumed same for both the groups. In the final step, steps 1-4 are repeated for all gait cycle instances to compute L^{kneeS} .

For all nine joint angle profiles, a joint abnormality index (i.e. AI^{joint}) is calculated by adding AI_i^{kneeS} at all instances (equation 5).

$$AI^{kneeS} = \sum_{i=1}^{51} AI_i^{kneeS} \quad (5)$$

These AI^{joint} are multiplied with their corresponding weight W^{joint} and summed to compute A-GAS.

$$A - GAS = \sum_{j=1}^9 AI^j * W^j \quad (6)$$

where, W^j is the weight of j th joint. Weight vector W represents prior probability of abnormality in joints. The values of W are derived from Table I.

Fig. 2. Top portion of the figure shows different events of the gait with respect to right leg (light blue). A: Mean and SD of knee sagittal joint angle profile for CP (red) and TD (blue) subjects is shown. X-axis is gait cycle instance (range 1 to 51) and Y-axis is joint angle in degree. B: PDFs of CP (red) and TD (blue) subjects at 30th gait cycle instance of knee sagittal joint angle profile. C: Plot of function used to convert P-value from T-test into Instance weight. D: Plot showing P-values (y-axis) and Instance weight vector for each gait cycle instance (x-axis).

TABLE I
NO OR MINOR GAIT DEVIATIONS PERCENT SAMPLE IN CP DATASET FOR VARIOUS JOINT ANGLE PROFILE ACCORDING TO RATINGS OF AN EXPERT [41].

Joint angle profile	No or minor gait deviations (%)
Knee Sagittal	23.04
Hip Sagittal	55.3
Hip Coronal	62.9
Hip Transverse	75.4
Ankle Sagittal	39.16
Pelvic Sagittal	16.0
Pelvic Coronal	48.6
Pelvic Transverse	44.4
Foot Transverse	66.6

The values from Table I are first assigned to W , then linearly mapped from 0.5 to 1 to and normalised by dividing $\sum W$. So, the final values of W become 0.941, 0.669, 0.605, 0.5, 0.805, 1, 0.726, 0.761 and 0.574. A similar reasoning is used for computing A-GAS by using only three joint angle profiles viz., ankle sagittal, knee sagittal and hip sagittal. Thus, the equation for A-GAS is given by the following equation 7.

$$A - GAS = \sum_{j=1}^3 AI^j * W^j \quad (7)$$

The final values of W becomes 1, 0.5 and 0.748. The pseudo-code of calculating A-GAS, AI and AII for a sample X is

given in Fig. 4. For all the analyses, a system having the following specifications is used; Intel(R) Core(TM) i5 6200U CPU @ 2.30 GHz, 4 GB DDR4 RAM, NVIDIA GeForce 930MX graphic engine, 64-bit Windows 10 Operating System and Matlab 2020a platform.

III. RESULTS

A. A sample report of A-GAS

By following the steps mentioned in the Section II, report of A-GAS is generated for a sample CP subject (see Fig. 5).

The report of A-GAS has four key features. First, it provides an easily interpretable comprehensive gait assessment. Second, it utilises the unique properties of each gait cycle instances to compute corresponding AII . Third, it provides abnormality indices at three levels for better understanding of gait pathology. Fourth, it uses knowledge of both CP and TD distributions to compute AII .

B. Comparison of A-GAS for CP and TD subjects

To understand the performance of A-GAS an analysis is performed. A-GAS is computed for all CP and TD subjects. Fig. 6 shows PDFs of CP and TD subjects. It can be interpreted from the Fig. 6 that A-GAS value of TD subjects is less than most of the CP subjects. The overlapping areas of PDFs is

Fig. 3. A: Knee sagittal joint angle profile of a sample X^{kneeS} (black), mean and SD of CP (red) subjects and mean and SD (blue) of TD subjects is shown. B: Position of X^{kneeS} in PDFs of CP (red) and TD (blue) subjects at 30th gait cycle instance of knee sagittal joint angle profile.

Input: $\mu^{TD}, \mu^{CP}, \sigma^{TD}, \sigma^{CP}, I, W$
Output: $AI, AI, A - GAS$
 Testing data: X
 for joint = {'kneeS', 'hipS', 'hipC', 'hipT', 'ankleS',
 'pelvicS', 'pelvicC', 'PelvicT', 'FootT'}:
 $AI^{joint} = 0$
 for i = 0: length(X):
 $LR_i^{joint} = t(\mu^{TD}, \sigma^{TD} | X_i^{joint}) / t(\mu^{CP}, \sigma^{CP} | X_i^{joint})$
 if $t(\mu^{CP}, \sigma^{CP} | X_i^{joint}) == 0$
 $L_i^{joint} = 0$
 else if $LR_i^{joint} < 1$
 $L_i^{joint} = 1 - LR_i^{joint}$
 else
 $L_i^{joint} = 0$
 end
 $AI_i^{joint} = L_i^{joint} * L_i^{joint}$
 $AI^{joint} = AI^{joint} + AI_i^{joint}$
 $A - GAS = A - GAS + AI^{joint}$

Fig. 4. The pseudo-code of calculating A-GAS, AI and AII for a sample X .

negligible, and by setting a decision criterion at around 0.13, A-GAS can accurately classify subjects into TD and CP.

A similar analysis is performed for individual joints. Fig. 7 shows PDFs of CP and TD subjects for all nine joints. It can be interpreted from the Fig. 7 that AI value of TD subjects is less than most of the CP subjects at all joints. However, there are noticeable overlapping of PDFs of AI for CP and TD subjects at hip coronal, pelvic coronal and pelvic transverse

joint angle profiles. There can be multiple possible reasons for these overlapping. One reason can be no or less deviation of CP gait angle profile from TD gait angle profile in many CP subjects at hip coronal, pelvic coronal and pelvic transverse joint angle profiles. This reasoning is also supported by results of [41], where 62.9% CP hip coronal angle profile, 48.6% pelvic coronal angle profile and 44.4% pelvic transverse angle profile showed no minor gait deviations. Another possible reason can be induction of minor abnormality in TD subjects at hip coronal, pelvic coronal and pelvic transverse joint angle profiles due to measurement errors. The range of motion of these joints is very small ($5^\circ - 20^\circ$), and small error in measurement can leads to a large deviation from normal gait angle profile.

C. Comparison of assessments

Establishment of a instrumented gait laboratory needs plenty of money and space. Moreover, recording and processing of 3-dimensional signal takes time [44]. However, low cost equipments such as Kinect, one-camera set-up and goniometer can be easily used for both small clinics and personal use. Therefore, a study is performed to propose a combination of three joint angle profiles that can replace current nine joint angle profile configuration effectively. A configuration having knee sagittal, hip sagittal and ankle sagittal joint angle profile is used for developing a three joint angle profile based score because its performance is found to be better than all other possible combinations. At all samples, Bland- Altman plot [45] is calculated for three joint angle profiles configuration (say A-GAS 1) and nine joint angle profiles configuration (say A-GAS 2). Fig. 8 shows, around 96% difference (A-GAS - 1 A-GAS 2) lies in 95% confidence level range. This suggests that both the configurations can be used interchangeably for clinical use [45].

D. Comparison of A-GAS with present rating based scale

As discussed in the Section I, most of the present scales are based on raters. Since all rating based scale has its pros and cons, it is very difficult to selected best scale among them. To address this problem, a consensus study was performed by Angela Nieuwenhuys et al. [46]. Thirteen experts from eight gait laboratories identified 49 clinical assessment relevant joint motion patterns in children with CP. Each trial of the CP dataset used in the study is rated by one of the two expert rater. '0'rating for a joint motion pattern denotes no or minor deviation from TD, while, rating other than '0'denotes a specific gait abnormality. This rater based method provides discrete value of rating while A-GAS provides continuous score. Therefore, for comparison all trials are divided into two classes; one with '0'rating and other with non-'0'ratings. Fig. 9 shows histograms of A-GAS for both the classes at each joint angle profile. In all the joint angle profiles, mean of A-GAS for class with '0'rating is less than mean of A-GAS for class with non-'0'rating. In a scenario, when both assessments are perfect we expect zero overlapping between the histograms with zero A-GAS value for all trials of class with '0'rating. However, in the Fig. 9 we can observe a lot of overlapping

Fig. 5. Report of A-GAS consists of nine subplots. Each subplot corresponds to one joint angle profile; A: knee sagittal, B: hip sagittal, C: hip coronal, D: hip transverse, E: ankle sagittal, F: pelvic sagittal, G: pelvic coronal, H: pelvic transverse and I: foot transverse. In a subplot, blue curve represents mean normal joint profile, red curve represents sample joint angle profile, height and color of bar plot represents magnitude of instance abnormality index (AII) at a instance, y-left axis represents angle value (in degree), y-right represents AII, x-axis represents gait cycle instance and numerical value at bottom of each subplot represents joint AI. A-GAS of the sample is written at the bottom of the plot.

Fig. 6. PDFs of A-GAS calculated for CP and TD samples.

between the histograms at some joint angle profiles. The possible reason for overlapping can be, incompleteness and insensitiveness of the rating based scale. Since, the rating based scale exploited a limited number of gait abnormal patterns to quantify abnormalities, some time it fails to capture abnormal

patterns. This limitation of rater based method contribute to its incompleteness to understand all possible types of gait abnormalities. Non-zero value of A-GAS value for trials of class with '0' rating supports this argument. Furthermore, the rater-based method is insensitive to minor changes in gait angle profile because it is based on a discrete scale. For example, same rating will be provided to a knee joint in sagittal with minor knee hyperextension and a knee joint in sagittal with major knee hyperextension [41]. While, A-GAS is sensitive to even minor changes as it provides continuous reading of gait abnormalities.

IV. DISCUSSION

Gait management in CP and other pathological gait disorders is a long process. It is crucial to track the progress of the process by timely recording patient's gait activities. However, rating a limited number of gait events with discrete values may lead to poor sensitivity of the method [26]. Moreover, during this long process of treatment, a patient may have to change their doctor. However, as most of the ways are subjective, the therapy may suffer after switching doctors. In addition, the subjectivity of these methods may lead to poor reliability, specificity, sensitivity, and validity [47]. Despite of many possible solutions of developing a robust automated gait assessment using deep learning based method, we choose a

Fig. 7. The figure consists of nine subplots. Each subplot corresponds to one joint angle profile; A: knee sagittal, B: hip sagittal, C: hip coronal, D: hip transverse, E: ankle sagittal, F: pelvic sagittal, G: pelvic coronal, H: pelvic transverse and I: foot transverse. In a subplot, blue curve represents PDF of joint abnormality index (AI) of TD sample for its corresponding joint angle profile and red curve represents PDF of AI of CP sample for its corresponding joint angle profile.

Fig. 8. Bland-Altman plot for three joint angle profiles configuration (say A-GAS 1) and nine joint angle profiles configuration (say A-GAS 2). (A-GAS - 1 A-GAS 2) difference for 1690 samples out of 1760 lies in 95% confidence level range.

conventional approach due to three reasons. First, it is very difficult to interpret/visualise the computations using by a deep neural network to predict gait assessment score. Conversely, a probabilistic approach based on simple computation provide sense to the intermediate computations. Such transparency in computations increases reliability and clinical acceptance of the method. Second, an optimum solution of a deep learning

Fig. 9. The figure consists of nine subplots. Each subplot corresponds to one joint angle profile; A: knee sagittal, B: hip sagittal, C: hip coronal, D: hip transverse, E: ankle sagittal, F: pelvic sagittal, G: pelvic coronal, H: pelvic transverse and I: foot transverse. In a subplot, blue region represents histogram of joint abnormality index (AI) of trials of class with '0' rating for its corresponding joint angle profile and red curve represents PDF of AI of trials of class with non-'0' rating for its corresponding joint angle profile.

based approaches corresponds to a local minima of the loss function. Use of the same network architecture on a similar problem do not guarantee the convergence of the solution. In other words, deep learning based approaches are problem and training data specific. However, simple tools of statistics and probability are problem and data invariant. A-GAS is a generalised approach that can also be used for other modalities of gait data such as gait kinetics. Third, implementation and understanding of deep learning based approaches require high performance computers (including GPUs/TPUs) and expertise in a specific field. While computations of A-GAS can be done on simple computers/smartphones. Also, elementary knowl-

edge of mathematics and programming is sufficient for its implement and understanding.

The study proposed an automated method for gait assessment with a great potential for clinical use. However, there are a few limitations of this study. the performance of A-GAS is only tested on CP a dataset, its performance on other gait pathologies needs to be tested. Also, the feasibility of A-GAS needs to be tested in clinical environments. Since the proposed method extracts decision parameters from the normal and abnormal data, its results depend on the population of the datasets. Moreover, as it compares gait cycle instances of a given sample with normal and abnormal datasets, the sampling frequency of the datasets should be matched with the sample. Furthermore, A-GAS can quantify abnormality in a lower limb, so it does not directly account for gait asymmetry. However, by comparing A-GAS, AI, and AII of both the legs, an index for abnormality can be derived. In future work, we will test A-GAS in a clinical setting; by understanding doctors' needs, some additional features will be added.

V. CONCLUSION

In conclusion, the study tries to improve the method of diagnosing gait abnormalities in CP by presenting an automated, comprehensive gait assessment score from objectively quantifying gait abnormalities in a lower limb. Performance of A-GAS is verified on Kinematic data of 356 CP and 41 TD children, which suggests great potential for clinical use. A graphical representation of the gait assessment report, including AII, AI, and A-GAS is generated to understand the results better. Moreover, the study compares the performance of A-GAS with a present rating-based gait assessment scores of the same dataset to understand fundamental differences between them. Lastly, the degree of agreement between values of A-GAS's for high cost multi camera set-up using nine joint angle profiles and values of A-GAS for low-cost single camera set-up using three joint angle profiles is computed. Results suggest both the set-ups can be used interchangeably, as there are no significant differences in performance of A-GAS for both the set-ups.

REFERENCES

- [1] W. Pirker and R. Katzenschlager, "Gait disorders in adults and the elderly," *Wiener Klinische Wochenschrift*, vol. 129, no. 3, pp. 81–95, 2017.
- [2] M. L. Aisen, D. Kerkovich, J. Mast, S. Mulroy, T. A. Wren, R. M. Kay, and S. A. Rethlefsen, "Cerebral palsy: clinical care and neurological rehabilitation," *The Lancet Neurology*, vol. 10, no. 9, pp. 844–852, 2011.
- [3] S. Armand, G. Decoulon, and A. Bonnefoy-Mazure, "Gait analysis in children with cerebral palsy," *EFORT open reviews*, vol. 1, no. 12, pp. 448–460, 2016.
- [4] A. A. Kareem, "Use of botulinum toxin in a cerebral palsy," *Cerebral Palsy-Clinical and Therapeutic Aspects*, pp. 95–112, 2018.
- [5] C. Rathinam, A. Bateman, J. Peirson, and J. Skinner, "Observational gait assessment tools in paediatrics—a systematic review," *Gait & posture*, vol. 40, no. 2, pp. 279–285, 2014.
- [6] D. H. Sutherland and J. R. Davids, "Common gait abnormalities of the knee in cerebral palsy," *Clinical orthopaedics and related research*, no. 288, pp. 139–147, 1993.
- [7] T. Winters, J. Gage, and R. Hicks, "Gait patterns in spastic hemiplegia in children and young adults," *J Bone Joint Surg Am*, vol. 69, no. 3, pp. 437–441, 1987.
- [8] J. Rodda and H. Graham, "Classification of gait patterns in spastic hemiplegia and spastic diplegia: a basis for a management algorithm," *European journal of neurology*, vol. 8, pp. 98–108, 2001.
- [9] T. F. Novacheck, J. L. Stout, and R. Tervo, "Reliability and validity of the gillette functional assessment questionnaire as an outcome measure in children with walking disabilities," *Journal of pediatric orthopaedics*, vol. 20, no. 1, p. 75, 2000.
- [10] H. K. Graham, A. Harvey, J. Rodda, G. R. Nattrass, and M. Pirpiris, "The functional mobility scale (fms)," *Journal of Pediatric Orthopaedics*, vol. 24, no. 5, pp. 514–520, 2004.
- [11] D. J. Russell, P. L. Rosenbaum, D. T. Cadman, C. Gowland, S. Hardy, and S. Jarvis, "The gross motor function measure: a means to evaluate the effects of physical therapy," *Developmental Medicine & Child Neurology*, vol. 31, no. 3, pp. 341–352, 1989.
- [12] J. W. Varni, T. M. Burwinkle, S. J. Berrin, S. A. Sherman, K. Artavia, V. L. Malcarne, and H. G. Chambers, "The pedsql in pediatric cerebral palsy: reliability, validity, and sensitivity of the generic core scales and cerebral palsy module," *Developmental medicine and child neurology*, vol. 48, no. 6, pp. 442–449, 2006.
- [13] S. M. Haley, W. J. Coster, Y.-C. Kao, H. M. Dumas, M. A. Fragala-Pinkham, J. M. Kramer, L. H. Ludlow, and R. Moed, "Lessons from use of the pediatric evaluation of disability inventory (pedi): Where do we go from here?" *Pediatric physical therapy: the official publication of the Section on Pediatrics of the American Physical Therapy Association*, vol. 22, no. 1, p. 69, 2010.
- [14] P. Thomason, A. Tan, A. Donnan, J. Rodda, H. K. Graham, and U. Narayanan, "The gait outcomes assessment list (goal): validation of a new assessment of gait function for children with cerebral palsy," *Developmental Medicine & Child Neurology*, vol. 60, no. 6, pp. 618–623, 2018.
- [15] N. L. Young, J. I. Williams, K. K. Yoshida, and J. G. Wright, "Measurement properties of the activities scale for kids," *Journal of clinical epidemiology*, vol. 53, no. 2, pp. 125–137, 2000.
- [16] H. S. Read, M. E. Hazlewood, S. J. Hillman, R. J. Prescott, and J. E. Robb, "Edinburgh visual gait score for use in cerebral palsy," *Journal of pediatric orthopaedics*, vol. 23, no. 3, pp. 296–301, 2003.
- [17] B. Toro, C. J. Nester, and P. C. Farren, "Inter-and intraobserver repeatability of the salford gait tool: an observation-based clinical gait assessment tool," *Archives of physical medicine and rehabilitation*, vol. 88, no. 3, pp. 328–332, 2007.
- [18] T. F. Novacheck, J. P. Trost, and S. Sohrweide, "Examination of the child with cerebral palsy," *Orthopedic Clinics of North America*, vol. 41, no. 4, pp. 469–488, 2010, orthopedic Management of Cerebral Palsy. [Online]. Available: <https://www.sciencedirect.com/science/article/pii/S0030589810000647>
- [19] K. G. Maathuis, C. P. van der Schans, A. Van Iperen, H. S. Rietman, and J. H. Geertzen, "Gait in children with cerebral palsy: observer reliability of physician rating scale and edinburgh visual gait analysis interval testing scale," *Journal of Pediatric Orthopaedics*, vol. 25, no. 3, pp. 268–272, 2005.
- [20] P. Araújo, R. Kirkwood, and E. Figueiredo, "Validity and intra-and inter-rater reliability of the observational gait scale for children with spastic cerebral palsy," *Brazilian Journal of Physical Therapy*, vol. 13, pp. 267–273, 2009.
- [21] T. A. Wren, S. A. Rethlefsen, B. S. Healy, K. P. Do, S. W. Dennis, and R. M. Kay, "Reliability and validity of visual assessments of gait using a modified physician rating scale for crouch and foot contact," *Journal of Pediatric Orthopaedics*, vol. 25, no. 5, pp. 646–650, 2005.
- [22] B. Toro, C. J. Nester, and P. C. Farren, "The development and validity of the salford gait tool: an observation-based clinical gait assessment tool," *Archives of physical medicine and rehabilitation*, vol. 88, no. 3, pp. 321–327, 2007.
- [23] E. Dugan, F. Trachtenberg, and M. A. Hall, "Development of abbreviated measures to assess patient trust in a physician, a health insurer, and the medical profession," *BMC health services research*, vol. 5, no. 1, pp. 1–7, 2005.
- [24] W. E. Dickens and M. F. Smith, "Validation of a visual gait assessment scale for children with hemiplegic cerebral palsy," *Gait & posture*, vol. 23, no. 1, pp. 78–82, 2006.
- [25] A. H. Mackey, G. L. Lobb, S. E. Walt, and N. S. Stott, "Reliability and validity of the observational gait scale in children with spastic diplegia," *Developmental medicine and child neurology*, vol. 45, no. 1, pp. 4–11, 2003.
- [26] M. Galli, V. Cimolin, C. Rigoldi, and G. Albertini, "Quantitative evaluation of the effects of ankle foot orthosis on gait in children with cerebral palsy using the gait profile score and gait variable scores,"

Journal of Developmental and Physical Disabilities, vol. 28, no. 3, pp. 367–379, 2016.

- [27] K. Hori, Y. Mao, Y. Ono, H. Ora, Y. Hirobe, H. Sawada, A. Inaba, S. Orimo, and Y. Miyake, "Inertial measurement unit-based estimation of foot trajectory for clinical gait analysis," *Frontiers in physiology*, vol. 10, p. 1530, 2020.
- [28] J. Gao, P. Gu, Q. Ren, J. Zhang, and X. Song, "Abnormal gait recognition algorithm based on lstm-cnn fusion network," *IEEE Access*, vol. 7, pp. 163 180–163 190, 2019.
- [29] T.-N. Nguyen and J. Meunier, "Applying adversarial auto-encoder for estimating human walking gait abnormality index," *Pattern Analysis and Applications*, vol. 22, no. 4, pp. 1597–1608, 2019.
- [30] T.-N. Nguyen, H.-H. Huynh, and J. Meunier, "Estimating skeleton-based gait abnormality index by sparse deep auto-encoder," in *2018 IEEE Seventh International Conference on Communications and Electronics (ICCE)*. IEEE, 2018, pp. 311–315.
- [31] X. Chen, S. Liao, S. Cao, D. Wu, and X. Zhang, "An acceleration-based gait assessment method for children with cerebral palsy," *Sensors*, vol. 17, no. 5, p. 1002, 2017.
- [32] C. Kirtley, *Clinical gait analysis: theory and practice*. Elsevier Health Sciences, 2006.
- [33] R. Baker, "Gait analysis methods in rehabilitation," *Journal of neuro-engineering and rehabilitation*, vol. 3, no. 1, pp. 1–10, 2006.
- [34] A. Assi, I. Ghanem, F. Lavaste, and W. Skalli, "Gait analysis in children and uncertainty assessment for davis protocol and gillette gait index," *Gait & Posture*, vol. 30, no. 1, pp. 22–26, 2009.
- [35] A. Muro-De-La-Herran, B. Garcia-Zapirain, and A. Mendez-Zorrilla, "Gait analysis methods: An overview of wearable and non-wearable systems, highlighting clinical applications," *Sensors*, vol. 14, no. 2, pp. 3362–3394, 2014.
- [36] K. J. Noonan, S. Halliday, R. Browne, S. O'Brien, K. Kayes, and J. Feinberg, "Interobserver variability of gait analysis in patients with cerebral palsy," *Journal of Pediatric Orthopaedics*, vol. 23, no. 3, pp. 279–287, 2003.
- [37] G. Gorton, D. Hebert, and B. Goode, "Assesment of the kinematic variability between 12 shriners motion analysis laboratories," in *6th Annual Meeting of the Gait and Clinical Movement Analysis Society, Sacramento, California, Gait and Posture*, 2001.
- [38] U. G. Narayanan, "The role of gait analysis in the orthopaedic management of ambulatory cerebral palsy," *Current opinion in pediatrics*, vol. 19, no. 1, pp. 38–43, 2007.
- [39] T. D. Laet, A. Nieuwenhuys, E. Papageorgiou, and K. Desloovere, "3D gait analysis data of children with CP used for gait classification," 4 2017. [Online]. Available: https://figshare.com/articles/dataset/3D_gait_analysis_data_of_children_with_CP_used_for_gait_classification/4877432
- [40] A. Nieuwenhuys, E. Papageorgiou, G. Molenaers, D. Monari, T. De Laet, and K. Desloovere, "Inter-and intrarater clinician agreement on joint motion patterns during gait in children with cerebral palsy," *Developmental Medicine & Child Neurology*, vol. 59, no. 7, pp. 750–755, 2017.
- [41] A. Nieuwenhuys, E. Papageorgiou, K. Desloovere, G. Molenaers, and T. De Laet, "Statistical parametric mapping to identify differences between consensus-based joint patterns during gait in children with cerebral palsy," *PLoS One*, vol. 12, no. 1, p. e0169834, 2017.
- [42] G. Bovi, M. Rabuffetti, P. Mazzoleni, and M. Ferrarin, "A multiple-task gait analysis approach: kinematic, kinetic and emg reference data for healthy young and adult subjects," *Gait & posture*, vol. 33, no. 1, pp. 6–13, 2011.
- [43] T. Lencioni, I. Carpinella, M. Rabuffetti, A. Marzegan, and M. Ferrarin, "Human kinematic, kinetic and emg data during different walking and stair ascending and descending tasks," *Scientific data*, vol. 6, no. 1, pp. 1–10, 2019.
- [44] R. Bajpai and D. Joshi, "Movenet: A deep neural network for joint profile prediction across variable walking speeds and slopes," *IEEE Transactions on Instrumentation and Measurement*, vol. 70, pp. 1–11, 2021.
- [45] C. Bunce, "Correlation, agreement, and bland–altman analysis: statistical analysis of method comparison studies," *American journal of ophthalmology*, vol. 148, no. 1, pp. 4–6, 2009.
- [46] A. Nieuwenhuys, S. Öunpuu, A. Van Campenhout, T. Theologis, J. De Cat, J. Stout, G. Molenaers, T. De Laet, and K. Desloovere, "Identification of joint patterns during gait in children with cerebral palsy: a delphi consensus study," *Developmental Medicine & Child Neurology*, vol. 58, no. 3, pp. 306–313, 2016.
- [47] B. Toro, C. Nester, and P. Farren, "A review of observational gait assessment in clinical practice," *Physiotherapy theory and practice*, vol. 19, no. 3, pp. 137–149, 2003.

Rishabh Bajpai is currently a PhD scholar at the Centre for Biomedical Engineering, Indian Institute of Technology (IIT) Delhi, and All India Institute of Medical Sciences (AIIMS) Delhi, India. He received his B.Tech degree in Mechanical Engineering from PDPM Indian Institute of Information Technology, Design and Manufacturing Jabalpur, Madhya Pradesh, India. He has been working in the area of Brain Computer Interface, Rehabilitation engineering, Neuroscience, Electroencephalography and development of biomedical instrumentation for

applications specific to assistive devices for disabled.

Deepak Joshi is a faculty at Biomedical Engineering at IIT Delhi and All India Institute of Medical Sciences Delhi. He received his PhD in biomedical engineering from Indian Institute of Technology (IIT) Delhi. He has been working in the area of neuroprosthetics and neurorehabilitation for last nearly fourteen years. Before joining to IIT Delhi, he worked at various places including National University of Singapore, Newcastle University UK, and University of Oregon, USA. Dr. Joshi's current research work combines experimental and compu-

tational techniques to understand the neural correlates during balancing and seamless transitions during walking. Besides that, he is actively engaged in projects related to development of wearable devices for applications specific to diagnosis of neuromuscular disorders, assistive devices for elderly and disabled, and biofeedback for rehabilitation in stroke patients.